

μALBUMIN IP CONTROL I

REF IMAL-0046



PITV-MALCon1-v11 (06/2021)

	Selectra PRO	Selectra MACH
mg/L	26.0 (19.2 – 32.8)	28.7 (21.2 – 36.2)
mg/dL	2.60 (1.92 - 3.28)	2.87 (2.12 - 3.62)

LOT 22-4012



2024-11


Preparation



Precautions

Place for future

Selectra MACH Barcode



www.elitechgroup.com



ELITech Clinical Systems SAS - Zone Industrielle - 61500 SEES FRANCE