

HDL LDL CALIBRATOR

REF HLCA-0041





FV-HLCA-v4 (11/2021)

HDL	LDL value for reagent lots before 21-0894 . <i>For newer lots refer to page 2</i>
35 mg/dL 0.91 mmol/L	60 mg/dL 1.56 mmol/L
Traceability : reference method approved by the CDC (Centers for Disease Control and Prevention)	

LOT 22-4018


2024-08



www.elitechgroup.com

LDL Reagent lots before 21-0894

selectra
PRO



HDL LDL Calibrator 0
S25 FV-HLCA
22-4018 2024-08

Place for future
Selectra MACH Barcode



HDL LDL CALIBRATOR

REF HLCA-0041





FV-HLCA-v4 (11/2021)

HDL	LDL value for reagent lots from 21-0894. For former lots refer to page 1
35 mg/dL 0.91 mmol/L	67 mg/dL 1.73 mmol/L
Traceability : reference method approved by the CDC (Centers for Disease Control and Prevention)	

LOT 22-4018

2024-08



www.elitechgroup.com

LDL Reagent lots from 21-0894

selectra
PRO



HDL LDL Calibrator 0
S25 FV-HLCA
22-4018 2024-08

Place for future
Selectra MACH Barcode

