



EU Declaration of Conformity



ELITechGroup B.V.
Van Rensselaerweg 4
6956 AV Spankeren
The Netherlands

declares under sole responsibility that the IVD medical device specified below (including the listed accessories) and to which this declaration relates, conforms to the provisions of:

- **Regulation (EU) 2017/746** of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU (“**IVDR**”)
- **Directive 2006/42/EC** of the European Parliament and of the Council of 17 May 2006 on machinery, and, amending Directive 95/16/EC (“**MD**”)
- **Directive 2011/65/EU** of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment, including Commission **Delegated Directive (EU) 2015/863** of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substances (“**RoHS**”).

This IVD medical device carries the CE-marking and is notified in accordance with the IVDR.

| Catalogue number | Description | GTIN |
|------------------|----------------|-----------------|
| 6004-301 | Selectra Mach5 | 3661540 60054 8 |

| | |
|-------------------------|--|
| Product | Clinical chemistry analyzer, automated |
| Basic UDI-DI | 3661540Mach-seriesTL |
| EMDN code | W02010101 |
| GMDN code | 56676 |
| Intended purpose | Automated clinical chemistry system intended for <i>in vitro</i> diagnostic measurements of analytes in samples derived from the human body. The system is intended for use in clinical laboratories and must be operated by qualified personnel. The system has a core module that consists of a spectrophotometric measurement system. |
| Risk Class | A, according to Rule 5 of Annex VIII |
| Product type | <input checked="" type="checkbox"/> <i>in vitro</i> diagnostic medical device <input type="checkbox"/> accessory for an <i>in vitro</i> diagnostic medical device |
| SRN | NL-MF-000021018 |

Spankeren, 28-Mar-2024

Adriaan Intveld

Person Responsible for Regulatory Compliance (PRRC)



EU Declaration of Conformity



List of applied (harmonized) standards

| | Standard version | Description | Tested / certified by |
|----------------------------------|--------------------------------|--|-----------------------|
| Safety | IEC 61010-1:2010/ AMD1:2016 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements | UL |
| | IEC 61010-2-010:2014 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of material | |
| | IEC 61010-2-051:2015 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-051: Particular requirements for laboratory equipment for mixing and stirring. | |
| | IEC 61010-2-101:2015 | Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment | |
| EMC | IEC 61326-1:2020 | Electrical equipment for measurement, control, and laboratory use - EMC requirements – Part 1: General requirements | DEKRA |
| | IEC 61326-2-6:2020 | Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-6: Particular requirements – In Vitro diagnostic (IVD) medical equipment | |
| Quality Management System | ISO 13485:2016 | Medical devices—Quality management systems—Requirements for regulatory purposes. | LRQA |