

	<h1>EU Declaration of Conformity</h1>	
---	---------------------------------------	---

**ELITechGroup B.V.**  
**Van Rensselaerweg 4**  
**6956 AV Spankeren**  
**The Netherlands**


declares under sole responsibility that the IVD medical device specified below and to which this declaration relates, conforms to the provisions of:

- **Regulation (EU) 2017/746** of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU ("IVDR")
- **Directive 2006/42/EC** of the European Parliament and of the Council of 17 May 2006 on machinery, and, amending Directive 95/16/EC ("MD")
- **Directive 2011/65/EU** of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment, including Commission **Delegated Directive (EU) 2015/863** of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substances ("RoHS").

This IVD medical device carries the CE-marking and is notified in accordance with the IVDR.

Catalogue number	Description	GTIN
6003-760	Viva-ProE	3661540 60035 7

Product	Clinical chemistry analyzer, automated
Basic UDI-DI	3661540Pro-series8A
EMDN code	W02010101
GMDN code	56676
Intended purpose	Automated clinical chemistry analyzer intended for use in clinical laboratories. It is intended to be used for a variety of assay methods that have been applied to spectrophotometric techniques.
Risk Class	A, according to Rule 5 of Annex VIII
Product type	<input checked="" type="checkbox"/> <i>in vitro</i> diagnostic medical device <input type="checkbox"/> accessory for an <i>in vitro</i> diagnostic medical device
SRN	NL-MF-000021018

Spankeren 27-Dec-2023,  Adriaan Intveld Person Responsible for Regulatory Compliance (PRRC)
---

### List of applied (harmonized) standards

	Standard version	Description	Tested / certified by
<b>Safety</b>	IEC 61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements	DEKRA
	IEC 61010-2-010:2014	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-010: Particular requirements for laboratory equipment for the heating of material	
	IEC 61010-2-081:2015	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes	
	IEC 61010-2-101:2015	Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment	
	UL 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1. General requirements	UL
<b>EMC</b>	IEC 61326-1:2020	Electrical equipment for measurement, control and laboratory use - EMC requirements – Part 1: General requirements	DEKRA
	IEC 61326-2-6:2020	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-6: Particular requirements – In Vitro diagnostic (IVD) medical equipment	
<b>Quality Management System</b>	ISO 13485:2016	Medical devices—Quality management systems—Requirements for regulatory purposes.	LRQA