

ELI.H.A *Aspergillus*

Serodiagnosis of aspergillosis by indirect haemagglutination

102 tests

(Ref. 44602)

**8000110-EN-2025-07**  
For *in vitro* diagnostic use only, for professional use only.  
Single use test.



**1 - AIM**  
ELI.H.A *Aspergillus* enables the quantitative determination of anti-*Aspergillus fumigatus* serum antibodies by indirect haemagglutination.  
Each kit allows 102 tests to be carried out or 17 reactions of 6 dilutions.

**2 - INTRODUCTION**  
*Aspergillus fumigatus* is the most frequently incriminated species in human pathology. *A. fumigatus* possesses a particular ability for parasitic adaptation in man.  
However, certain favourable conditions are necessary for its development:  
- local conditions (preformed cavities, debrided pulmonary abscesses, impaired mucous agglutinins from the serum by adsorption, membranes...);  
- general conditions (immunosuppression following major surgery or medical procedures: organ transplantation, immunosuppressive therapy, steroids, antibiotics...).

**3 - PRINCIPLE**  
ELI.H.A *Aspergillus* is based on the indirect haemagglutination principle. The sensitized red blood cells consist of sheep red blood cells covered with an *Aspergillus fumigatus* antigen.  
The presence of specific serum antibodies results in agglutination of the sensitized red blood cells resulting in a cloudy red/brown deposit coating the well. In the absence of specific antibodies, the red blood cells form a ring-like deposit at the bottom of the well.  
The non-sensitized red blood cells ensure the specificity of the reaction making it possible to eliminate any interference from the natural anti-sheep agglutinins (Forssman heteroantibodies, infectious mononucleosis antibodies...).  
The reaction is carried out in a U-microplate.  
Handling is simple and fast, with results within 2 hours.

**4 - REAGENTS AND MATERIAL**

Description	Quantity
<b>R1:</b> Vial of 2,2 mL of sensitized red blood cells	1
<b>R2:</b> Vial of 1 mL of non-sensitized red blood cells	1
<b>BUF:</b> Vial of 55 mL of phosphate buffer pH 7.2	1
<b>R3:</b> Vial of 2 mL of adsorbent	1
<b>CONTROL +:</b> Vial of 0,2 mL of titrated positive control	1
<b>CONTROL -:</b> Vial of 0,2 mL of negative control	1
<b>MICROPLATE:</b> Microplate with a U-bottom	2
<b>DROPPER:</b> Special dropper	2

**5 - PRECAUTIONS**

- The reagents are intended for *in vitro* diagnostic use only and must be handled by authorized personnel.
- Tests are for single use only.
- All the reagents, except the **BUF** reagent, contain raw materials of animal origin and must be handled with caution.
- Patient samples are potentially infectious. They must be handled with caution, in observance of hygiene rules and the current regulations for this type of product in the country of use.
- The reagents contain sodium azide (concentration < 0.1%). **The sodium azide contained in the reagents can react with the heavy metals in the pipes to form explosive compounds. It is therefore recommended not to dispose of the reagents down the sink and to follow the recommendations and regulations for waste disposal in force.**
- Do not use reagents after the expiry date.
- Do not use reagents from different batch numbers.
- Prior to use, allow the serum and the reagents to reach room temperature.
- Carefully shake the **R1** and **R2** reagents before use.
- When dispensing the **R1** and **R2** reagents, make sure that the dropper is perfectly vertical. Check for the absence of air bubbles in the drops to ensure constant delivery volumes.

**6 - SAMPLE COLLECTION AND TREATMENT**  
Use fresh serum or serum preserved at -20°C, and not showing any sign of haemolysis, cloudiness or of contamination.  
Avoid repeated freezing and defrosting.  
Do not decompliment the serum.

**7 - STABILITY, STORAGE AND PREPARATION OF REAGENTS**  
The reagents are ready-to-use.  
All the reagents stored at 2-8°C, in their original packaging, are stable until the expiry date indicated on the box. Do not freeze.

**8 - MATERIAL REQUIRED BUT NOT SUPPLIED**

- Automatic pipette(s) with a pipetting volume adapted to the volume that will be measured;
- Contaminated waste containers;
- Centrifuge;
- Haemolysis tubes.

**9 - METHOD**  
Allow the reagents to reach room temperature before use.

**9.1 - Sample preparation**  
Carry out a 1:40 dilution of the serum to be tested:

- 50 µL of serum ;
- 1,95 mL of **BUF** reagent.

**9.2 - Realization of the test on a microplate**

- Using a multichannel micropipette, add 50 µL of **BUF** reagent to 8 wells of the microplate.
- Using a micropipette, add 50 µL of diluted serum to the 1 well.  
Mix the serum with the **BUF** reagent and carry out a serial dilution, preferably using a microdiluter, by transferring 50 µL from the 1st well into the 2nd, then 50 µL from the 2nd to the 3rd, and so on until the 6th well is reached. 50 µL from the 6th well is then discarded.  
In this way, dilutions from 1:80 to 1:2560 are obtained.
- Add 50 µL of diluted serum to the 7th well.  
Mix the serum with the **BUF** reagent and then discard 50 µL.  
This dilution (1:80) is the serum control, whose role is to detect the natural anti-sheep agglutinins that could be present in certain serum samples.
- Carefully shake the **R1** and **R2** reagents.
  - Add 1 drop of **R1** reagent to the first 6 wells.
  - Add 1 drop of **R2** reagent to the 7th well (serum control).
  - Add 1 drop of **R1** reagent to the 8th well (reagent control) whose role is to control the validity of the **BUF** and **R1** reagent.

**Note:** Only carry out one reagent control for each series of tests.

- Very carefully, shake the contents of the wells:
  - either manually, by tapping laterally the side of the microplate that has been posed flat on the bench;
  - or by using a vibrating plate shaker for microtiter plates (for example at 1300 rpm for 10 seconds). Do not use an orbital shaker.
- Now leave the plate to rest, away from any sources of vibration.
- The plate can be read after 2 hours.


**9.3 - Adsorption of the natural anti-sheep agglutinins in the event of agglutination of the serum control**

- Carefully shake the **R3** reagent.
- In a tube, add and mix:
  - 0,1 mL of serum ;
  - 0,3 mL of **R3** reagent.
- Incubate at room temperature for 60 minutes.
- Centrifuge at 2000 rpm for 15 minutes.
- Collect the supernatant; the serum is now at a 1:4 dilution
- Carry out a 1:10 dilution of the supernatant in **BUF** reagent to obtain an adsorbed stock dilution (1:40).
- Follow the steps described in "Realization of the test on a microplate", but replace the stock dilution by the adsorbed stock dilution.

**10 - READING**

**Negative reaction:** **Absence of haemagglutination.**  
Presence of a more or less large ring at the bottom of the well.

**Positive reaction:** **Presence of haemagglutination.**  
Presence of a cloudy red/brown deposit coating the well, sometimes there is the presence of a fine peripheral border.



**11 - INTERPRETATION OF RESULTS**

**Titer < 1/320:** **Non significant reaction.**  
Probable absence of a deep aspergillosis infection.  
Renew the test 2 to 3 weeks later and also carry out an electrosyneresis or an immunoelectrophoresis test.

**Titer = 1/320:** **Doubtful reaction.**  
Renew the test 2 to 3 weeks later and also carry out an electrosyneresis or an immunoelectrophoresis test.

**Titer ≥ 1/640:** **Significant reaction in favour of a deep aspergillosis infection.**

**12 - INTERNAL QUALITY CONTROL**  
The **CONTROL +** and **CONTROL -** reagents must be treated like test serums. The titer of the **CONTROL +** reagent must be the same as the titer printed on the vial label ± one dilution. There must not be any haemagglutination of the **CONTROL -**. If haemagglutination is present then the test is not valid.

**13 - CAUSES OF ERROR AND TEST LIMITS**

- Poor conservation of the serum.
- Poor conservation of the reagents after opening.
- Only use the droppers provided in the kit.
- Do not interchange the droppers between the **R1** and **R2** reagents.
- In the case of a positive reaction in the first 6 wells, carry out a further serial dilution in order to determine the titer limit of haemagglutination.
- The serum control must give a negative reaction (ring). In the event of haemagglutination of this control, it will be necessary to renew the test after having eliminated the natural anti-sheep agglutinins from the serum by adsorption.
- The reagent control must give a negative reaction (ring). In the event of haemagglutination of this control, the **ELI.H.A Aspergillus** cannot be used.
- Certain serums, whose antibody concentration is very high, can give rise to a zone phenomenon (with disappearance of the clouding) in the initial dilutions, which disappears in the subsequent dilutions.
- The quality of the reagents makes it possible to carry out the reaction in the evening and to read the test the following morning, provided that the microplate is not moved in any way and is protected from any sources of vibration.
- In all cases, it is necessary that the clinical, epidemiologic and biological data are taken fully into consideration before establishing the final diagnosis

**14 - PERFORMANCE**  
**ELI.H.A Aspergillus** consists of red blood cells sensitized by an *Aspergillus fumigatus* antigen, that ensures the specificity and sensitivity of the indirect haemagglutination reaction.  
Evaluation has demonstrated that the test has a sensitivity of 80% and a specificity of 98 %.

**15 - WASTE ELIMINATION**  
Waste should be disposed of in accordance with the hygiene rules and current regulations for this kind of product in the country of use.  
If the **BUF** reagent is spilled, clean the work area with absorbent paper and rinse with water. If a serum or another reagent is spilled on the work area, clean using bleach and absorbent paper.

**16 - BIBLIOGRAPHY**

1. J. CAPDEVILLE, S. FIABANE - Diagnostic immunologique des aspergilloses - *Le Pharmacien Biologiste*, Tome XII, N° 121, 205-249.
2. B. PESSON, N. LEGER, G. MADULO-LEBLOND - Diagnostic immunologique en parasitologie et en mycologie - *Le Pharmacien Biologiste*, Tome XIII, N° 123, 417-455.
3. J.-M. SENET, C. BRISSET - The diagnosis of aspergillosis by passive haemagglutination - *Biomedicine*, 1973, 19, 365-368.
4. J.-M. SENET, R. ROBERT, E. PICHOT - Intérêt de l'hémagglutination indirecte dans le diagnostic précoce de l'aspergillose - *Bulletin de la Société de Mycologie Médicale*, 1974, Tome III, N° 1, 45-48.
5. J.-M. SENET, R. ROBERT - Intérêt de l'hémagglutination dans le diagnostic de maladies parasitaires. Application à la toxoplasmose et à l'aspergillose - *Archives Médicales de l'Ouest*, 1979, Tome 11, N° 1, 39-42.
6. A. CULINO, M. MIEGEVILLE, O. MORIN - Apports de l'hémagglutination indirecte dans le diagnostic de l'aspergillose pulmonaire - *Feuillets de Biologie*, 1984, Vol. XXV, N° 137, 51/57.
7. M. MIEGEVILLE, O. MORIN, C. VERMEIL - Etude sérologique rétrospective concernant une série d'aspergillose chez des malades à "hauts risques" - *Feuillets de Biologie*, 1986, Vol. XXVII, N° 148, 37-39.
8. M.-C. GBADAMASSI, C. CIVADIER, T. SANDRE, T.-H. DUONG, C. COMBESCOT - Diagnostic immunologique de l'aspergillose : Proposition de l'association de deux techniques facilement mises en œuvre au laboratoire - *Feuillets de Biologie*, 1989, Vol. XXX, N° 169, 31-34.

The changes from the previous version are highlighted in grey.